



# Certificate of Calibration

Model: 02011

Serial No.: DHN001

Wrench Capacity: 2-24NM

Inspector:

Date of Calibration: 2020 03 27

Unit: N.W

Set Point	(CW) $\pm 4\%$		Actual Readings		
	Min	Max			
4	4.2	3.8	4.2	4.1	3.9
15	15.6	14.4	15.5	15.3	14.9
20	20.8	19.2	20.6	20.5	20
24	24.9	23	24.6	24.2	23.5

Testing Standard No.: DIN ISO 6789

Testing is in compliance with international Standards with test equipment and reference standards calibrated by a laboratory traceable to international standards.

Tester Model: T-50

Serial No.: DHN001

The uncertainty of the test equipment at 20°C is  $< \pm 1\%$

# EXAMPLE